



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CALIBRATOR LOW
Product Code : 226401
Lot No. : F2001421
Expiration date : 2023-04-18

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	F201401421	2023-04-18
CAL II	4 vials	F201401421	2023-04-18
CAL III	4 vials	F201401421	2023-04-18

Analytical Data	Specification	Result
Concentration	CAL I : < 15 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 ng/ml 50 ng/ml 109 ng/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-11-26

Quality Control Manager :

S.LECOURT

