



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN AT

Item Code : 120002

Lot No. : F2001540

Expiration date : 2023-05-27

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	2 vials	F204K01540	2023-05-27
R2 : Reaction Buffer	2 vials	F204K01540	2023-05-27

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	$A_{620} (0\%) \leq 0.15$ $A_{620} (0-150\%) \geq 0.40$ $R^2 \geq 0.98$ Controls measured within acceptance ranges	0.091 0.41 ≥ 0.98 Controls measured within acceptance range
Detection threshold	$\leq 10\%$	8%
Hook effect	$>200\%$	300%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-12-18

Quality Control Manager : **S.LECOURT**

