



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description: LIAPHEN Fibrinogen

Item Code: 120102

Lot No.: F2001457

Expiration date: 2023-05-07

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	F201901457	2023-05-07

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	Delta A620 (0-20µg/ml) \geq 0.40 Controls measured within acceptance ranges $R^2 \geq$ 0.98	0.714 Controls are in acceptance ranges > 0.99
Limit of Quantification	\leq 1 µg/mL (in the test dilution (1:200))	0.25 µg/mL
Hook effect	> 90 µg/mL (in the test dilution (1:200))	150 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-11-30

Quality Control Manager: S. LECOURT