



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

**CERTIFICATE OF ANALYSIS**

**Product Description:** BIOPHEN ARIXTRA CALIBRATOR  
**Product Code:** 222501  
**Lot No.:** F2001542  
**Expiration date:** 2023-05-16

Components	Quantity	Lot	Expiration Date
CAL 1	3 vials	F201501542	2023-05-16
CAL 2	3 vials	F201501542	2023-05-16
CAL 3	3 vials	F201501542	2023-05-16
CAL 4	3 vials	F201501542	2023-05-16

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 0.05 \mu\text{g/mL}$ CAL 2 : $0.40 - 0.60 \mu\text{g/mL}$ CAL 3 : $0.90 - 1.10 \mu\text{g/mL}$ CAL 4 : $1.35 - 1.60 \mu\text{g/mL}$	0 0.52 1.02 1.49
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	$R^2 = 1.00$ Measured values for controls in acceptance ranges

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2020-12-14

Quality Control Manager: S. LECOURT