



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN ORGARAN CONTROL
Product Code : 223501
Lot No. : F2001607
Expiration date : 2023-05-23

Components	Quantity	Lot	Expiration Date
C1	6 vials	F201701607	2023-05-23
C2	6 vials	F201701607	2023-05-23

Analytical Data	Specification	Result
Concentration	C1 : 0.40 -0.60 U/ml C2 : 0.80 -1.20 U/ml	0.49 0.99

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-12-17

Quality Control Manager : S.LECOURT