



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN ARIXTRA CONTROL PLASMA
Product Code: 224001
Lot No.: F2001541
Expiration date: 2023-05-16

Components	Quantity	Lot	Expiration Date
C1	6 vials	F201601541	2023-05-16
C2	6 vials	F201601541	2023-05-16

Analytical Data	Specification	Result
Concentration	C1 : 0.30 -0.50 µg/ml C2: 1.05 -1.35 µg/ml	0.43 1.23

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-12-14

Quality Control Manager: S. LECOURT