



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN FVIIa CONTROL SET
Product Code: 224901
Lot No.: F2001547
Expiration date: 2023-05-12

Components	Quantity	Lot	Expiration Date
C1	3 vials	F201501547	2023-05-12
C2	3 vials	F201501547	2023-05-12

Analytical Data	Specification	Result
Concentration	C1 : 50 -100 mIU/mL C2 : 200 -300 mIU/mL	75 mIU/mL 252 mIU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-12-14

Quality Control Manager: S. LECOURT