



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CONTROL
Product Code : 225301
Lot No. : F2001545
Expiration date : 2023-05-15

Components	Quantity	Lot	Expiration Date
C1	6 vials	F201A01545	2023-05-15
C2	6 vials	F201A01545	2023-05-15

Analytical Data	Specification	Result
Concentration	C1 : 140 -260 ng/ml C2 : 340 -460 ng/ml	187 ng/mL 408 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-12-21

Quality Control Manager : S.LECOURT

