



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN APIXABAN CALIBRATOR
Product Code: 226201
Lot No.: F2001513
Expiration date: 2023-05-09

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	F201601513	2023-05-09
CAL 2	4 vials	F201601513	2023-05-09
CAL 3	4 vials	F201601513	2023-05-09

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/mL CAL 2 : 200 -400 ng/mL CAL 3 : 500 -700 ng/mL	0 ng/mL 299 ng/mL 601 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	$R^2 = 0.99$ Measured values for controls in acceptance ranges

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-12-14

Quality Control Manager: S. LECOURT

