



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: COLLAGEN
Product Code: AG005K
Lot No.: F2001543
Expiration date: 2023-05-12

Components	Quantity	Lot	Expiration Date
COLLAGEN REAGENT	3 vials	F2001543	2023-05-13
COLLAGEN DILUENT	3 vials	F2001543	2023-05-12

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 10 µg/ mL	≥ 70%	96%
Reactivity (% AgregMax) at 2 µg/mL	≥ 70%	84%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-12-08

Quality Control Manager: S. LECOURT

