



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :**      **CALCIUM CHLORIDE 0.025M**  
**Product Code :**              **AR001K**  
**Lot No. :**                        **F2001714**  
**Expiration date :**              **2023-06-09**

Components	Quantity	Lot	Expiration Date
CALCIUM CHLORIDE 0.025M	12 vials (AR001K)	F204X01714	2023-06-09

Analytical Data	Specification	Result
aPTT on citrated human plasma (CEPHEN)	ratio $\leq$ 1.20	0.93
FVIII:C clotting assay	Delta CT (100%) with control sample $\leq$ 5%	0.3

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-01-19

Quality Control Manager : S.LECOURT