



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF: CBA

Product Code : RK038A

Lot No. : F2000818

Expiration date : 2023-06-27

Components	Quantity	Lot	Expiration Date
Collagen pre-coated plate	12x8 wells	F2000955	2024-01-03
vWF standard	3 vials	F2000986	2024-01-08
vWF Control I	1 vial	F2000988	2024-01-08
vWF Control II	1 vial	F2000989	2024-01-08
Anti-(h)- vWF HRP immunoconjugate	3 vials	F2000957	2024-01-15
Sample diluent	2x50 ml	F2000652	2023-10-30
Wash solution	1x50 ml	F2000037	2023-06-27
Conjugate diluent	1x25 ml	F2000351	2023-09-26
TMB	1x25 ml	190705D01	2023-07-30
Sulfuric Acid 0,45M	1x6 ml	F2000076	2023-07-04

Analytical Data	Specification	Result
Reactivity	A450 for 100 % \geq 1.00	1.88
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	Controls measured values within acceptance ranges
Blank	< 0.100	0.005
Detection threshold	\leq 5%	< 1%

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.052	1.434	0.882	0.403	0.213

	Acceptance Range	MV
CI	[77 - 104]	96
CII	[34 - 52]	44

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-08-24

Quality Control Manager : S.LECOURT

