



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Item Description : LIAPHEN Fibrinogen	
Item Code : 120102	
Lot No. : F2001731	
Expiration date : 2023-07-08	

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	F201B01731	2023-07-08

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	Delta A620 (0-20µg/ml) ≥ 0.40 Controls measured within acceptance ranges R ² ≥ 0.98	0.651 Controls are in acceptance ranges 1.00
Limit of Quantification	≤ 1 µg/ml (in the test dilution (1:200))	0.25 µg/mL
Hook effect	> 90 µg/ml (in the test dilution (1:200))	130 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-01-21

Quality Control Manager : S.LECOURT