



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR

Product Code : 222701

Lot No. : FA037217/FA037314

Expiration date : 2023-08-29

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA0372BA17/FA037314	2023-08-29
CAL 2	4 vials	FA0372BB17/FA037314	2023-08-29
CAL 3	4 vials	FA0372BC17/FA037314	2023-08-29

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 262 526
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-03-26

Quality Control Manager : S.LECOURT