



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CONTROL LOW
Product Code : 225101
Lot No. : FA037519/FA037613
Expiration date : 2023-08-29

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA0375DM19/FA037613	2023-08-29
C2	6 vials	FA0375DN19/FA037613	2023-08-29

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 45 -105 ng/ml	24 80

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-03-26

Quality Control Manager : S.LECOURT