



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR LOW

Product Code : 226001

Lot No. : FA037417

Expiration date : 2023-08-01

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FA0374DJ17	2023-08-01
CAL II	4 vials	FA0374DK17	2023-08-01
CAL III	4 vials	FA0374DL17	2023-08-01

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 52 102
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	Pass Pass

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-03-26

Quality Control Manager : S.LECOURT

