



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CALIBRATOR

Product Code : 226201

Lot No. : FA043918

Expiration date : 2023-08-07

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA0439DO18	2023-08-07
CAL 2	4 vials	FA0439DP18	2023-08-07
CAL 3	4 vials	FA0439DQ18	2023-08-07

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -400 ng/ml CAL 3 : 500 -700 ng/ml	0 304 610
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	Pass

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-03-10

Quality Control Manager : S.LECOURT

