



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CALIBRATOR
Product Code : 226501
Lot No. : FA044117
Expiration date : 2023-08-07

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA0441aa17	2023-08-07
CAL 2	4 vials	FA0441ab17	2023-08-07
CAL 3	4 vials	FA0441ac17	2023-08-07

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 40 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 253 497
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	Pass

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-03-10

Quality Control Manager : S.LECOURT

