



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CONTROL PLASMA
Product Code : 224501
Lot No. : FA065719
Expiration date : 2023-09-09

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA0657BD19	2023-09-09
C2	6 vials	FA0657BE19	2023-09-09

Analytical Data	Specification	Result
Concentration	C1 : 50 -150 ng/ml C2 : 250 -350 ng/ml	97 ng/mL 314 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-05-26

Quality Control Manager :

S.LECOURT