

CERTIFICATE OF ANALYSIS

Product Description :	PLASMA ARGATROBAN CALIBRATOR
Product Code :	SC030K
Lot No. :	FA062415
Expiration date :	2023-09-09

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA0624BZ15	2023-09-09
CAL 2	4 vials	FA0624CA15	2023-09-09
CAL 3	4 vials	FA0624CB15	2023-09-09
CAL 4	4 vials	FA0624CC15	2023-09-09
CAL 5	4 vials	FA0624CD15	2023-09-09

Analytical Data	Specification	Result
CAL 1	≤ 0.05 µg/mL	0
CAL 2	0.35 - 0.65 µg/mL	0.59
CAL 3	0.85 - 1.15 µg/mL	1.06
CAL 4	1.30 -1.65 µg/mL	1.60
CAL 5	1.80 -2.10 µg/mL	2.02

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-04-09

Quality Control Manager : S.LECOURT

