

CERTIFICATE OF ANALYSIS

Product Description : ARGATROBAN CONTROL PLASMA
Product Code : SC035K
Lot No. : FA062515
Expiration date : 2023-09-09

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA0625CE15	2023-09-09
C2	6 vials	FA0625CF15	2023-09-09

Analytical Data	Specification	Result
C1	0.55 - 0.80 µg/mL	0.69
C2	1.10 - 1.40 µg/mL	1.29

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-04-09

Quality Control Manager : S.LECOURT

