



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN a2 ANTIPLASMIN (LRT)
Product Code : 220502
Lot No. : FA20071C
Expiration date : 2023-10-15

Components	Quantity	Lot	Expiration Date
R1 :Plasmin	3 vials	FA2007171C	2023-10-15
R2 : SPm-41 Substrate	3 vials	FA2007181C	2023-10-15

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	Delta A405(100-50%)> 0.05 A405/min(100%) = 0.25 ± 0.10 R ² ≥ 0.98 Measured values for controls in acceptance ranges	0.089 0.25 PASS PASS
Detection threshold (LOQ) on CS	≤ 10%	< 5%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-11-16

Quality Control Manager : S.LECOURT