



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN PROTEIN C 5
Product Code : 221205
Lot No. : FA083327
Expiration date : 2023-10-09

Components	Quantity	Lot	Expiration Date
R1 : Protac	4 vials	FA0833FJ27	2023-10-16
R2 : PC substrate	4 vials	FA0833FK27	2023-10-09

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	A405(100%) \geq 0.80 A405(0%) \leq 0.10 R ² \geq 0.98 Measured values for controls in acceptance ranges	1.024 0.013 1.000 PASS
Detection Threshold	\leq 5%	< 2 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-05-10

Quality Control Manager : S.LECOURT