



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN ORGARAN CALIBRATOR  
**Product Code :** 222201  
**Lot No. :** FA083816  
**Expiration date :** 2023-10-09

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA0838CG16	2023-10-09
CAL 2	4 vials	FA0838CH16	2023-10-09
CAL 3	4 vials	FA0838CI16	2023-10-09
CAL 4	4 vials	FA0838CJ16	2023-10-09
CAL 5	4 vials	FA0838CK16	2023-10-09

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 0.05$ U/ml CAL 2 : 0.30 -0.50 U/ml CAL 3 : 0.70 -0.90 U/ml CAL 4 : 1.10 -1.30 U/ml CAL 5 : 1.45 -1.75 U/ml	0 U/mL 0.39 U/mL 0.78 U/mL 1.18 U/mL 1.54 U/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-05-17

Quality Control Manager : S.LECOURT

F820-11/BI/2201 -v10