



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN LMWH CONTROL LOW  
**Product Code :** 223701  
**Lot No. :** FA1005  
**Expiration date :** 2023-10-07

Components	Quantity	Lot	Expiration Date
CI	6 vials	FA0834BQ1C	2023-10-07
CII	6 vials	FA0835BR1A	2023-10-07

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml CII : 0.35 -0.65 IU/ml	0.28 0.52

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-05-10

Quality Control Manager : S.LECOURT