



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN FXIa CONTROL SET
Product Code : 224801
Lot No. : FA092416
Expiration date : 2023-10-20

Components	Quantity	Lot	Expiration Date
C1	3 vials	FA092416	2023-10-20
C2	3 vials	FA092416	2023-10-20

Analytical Data	Specification	Result
Concentration	C1 : about 10 mIU/ml C2 : about 30 mIU/ml	11 mIU/mL 27 mIU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-05-31

Quality Control Manager : S.LECOURT