



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST DDimer

Product Code : RK023A

Lot No. : FA1360

Expiration date : 2023-10-30

Components	Quantity	Lot	Expiration Date
Anti-(h) DDimer pre-coated plate	12x8 wells	FA1365	2024-01-07
DDimer standard	3 vials	FA1361	2023-12-29
Control I	1 vial	FA1362	2023-12-29
Control II	1 vial	FA1363	2023-12-29
Anti-(h)- DDimer HRP immunoconjugate	3 vials	FA1364	2024-01-12
Sample diluent	2x50 ml	F2000654	2023-10-30
Wash solution	1x50 ml	FA0257	2024-07-03
Conjugate diluent	1x25 ml	F2001562	2024-05-08
TMB	1x25 ml	200204D01	2024-02-28
Sulfuric Acid 0,45M	1x6 ml	F2001723	2024-06-19

Analytical Data	Specification	Result
Reactivity	A450 for 185 ng/ml \geq 1.40	1.65
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.004
Detection threshold	\leq 2 ng/ml	< 1

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 23.08.2021

Quality Control Manager : S.LECOURT