



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN ARIXTRA CALIBRATOR
Product Code : 222501
Lot No. : FA105417
Expiration date : 2023-11-14

Components	Quantity	Lot	Expiration Date
CAL 1	3 vials	FA1054CN17	2023-11-14
CAL 2	3 vials	FA1054CO17	2023-11-14
CAL 3	3 vials	FA1054CP17	2023-11-14
CAL 4	3 vials	FA1054CQ17	2023-11-14

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 0.05 \mu\text{g/ml}$ CAL 2 : $0.40 - 0.60 \mu\text{g/ml}$ CAL 3 : $0.90 - 1.10 \mu\text{g/ml}$ CAL 4 : $1.35 - 1.60 \mu\text{g/ml}$	0.00 $\mu\text{g/mL}$ 0.51 $\mu\text{g/mL}$ 1.01 $\mu\text{g/mL}$ 1.49 $\mu\text{g/mL}$
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	$R^2 > 0.99$ Measured values for controls in acceptance ranges

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-06-09

Quality Control Manager : S.LECOURT

