



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN APIXABAN CONTROL LOW  
**Product Code :** 225201  
**Lot No. :** FA11161A  
**Expiration date :** 2023-11-28

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA1116DW1A	2023-11-28
C2	6 vials	FA1116DX1A	2023-11-28

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 45 -105 ng/ml	25 ng/mL 81 ng/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-06-18

Quality Control Manager : S.LECOURT