



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN EDOXABAN CALIBRATOR  
**Product Code :** 226501  
**Lot No. :** FA109818  
**Expiration date :** 2023-11-21

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA1098aa18	2023-11-21
CAL 2	4 vials	FA1098ab18	2023-11-21
CAL 3	4 vials	FA1098ac18	2023-11-21

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 40$ ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/mL 242 ng/mL 489 ng/ mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-06-21

Quality Control Manager : S.LECOURT

