



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	Factor XI Deficient Plasma
Product Code :	DP070A/K
Lot No. :	FA111318
Expiration date :	2023-11-27

Components	Quantity	Lot	Expiration Date
Factor XI Deficient Plasma	1 vial (DP070A) 6 vials (DP070K)	FA1113JX18	2023-11-27

Analytical Data	Specification	Result
Calibration curve	CT (0%) > 150 sec CT (100%) < 100 sec	MC10: 155 sec STAR: 144 sec (*) MC10 : 58 sec STAR : 58 sec
Concentration FXI	< 1%	< 1%
Concentration FV	> 50 %	80%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

(*) Accepted under waiver (NCI 1926)

Date of Approval : 2023-06-23

Quality Control Manager : S.LECOURT

