



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

**Product Description : ZYMUTEST PF4**

**Product Code : RK006A**

**Lot No. : FA0716**

**Expiration date : 2023-11-17**

Components	Quantity	Lot	Expiration Date
PF4 pre-coated plate	12x8 wells	FA0730	2023-11-19
PF4 Standard	3 vials	FA0732	2023-11-24
PF4 Control I	1 vial	FA0733	2023-11-24
PF4 Control II	1 vial	FA0734	2023-11-24
PF4 HRP immunoconjugate	3 vials	FA0731	2023-11-27
PF4 Sample diluent	2x50 ml	FA0729	2023-11-17
Wash solution	1x50 ml	F2000802	2023-11-20
Conjugate diluent	1x25 ml	F2001562	2024-05-08
TMB	1x25 ml	200204D01	2024-02-28
Sulfuric Acid 0,45M	1x6 ml	F2001723	2024-06-19

Analytical Data	Specification	Result
Reactivity	A450 for 8 IU/ml $\geq 1.70$	1.63 (*)
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	$< 0.100$	0.06
Detection threshold	$\leq 0.5$ IU/ml	$< 0.1$

(\*) Accepted under waiver NCI N°1928

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-06-24

Quality Control Manager : S.LECOURT

