



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS
Product Description : ZYMUTEST TOTAL TAFI-Ag
Product Code : RK008A
Lot No. : FA1741
Expiration date : 2023-11-20

Components	Quantity	Lot	Expiration Date
Anti-(h) Total TAFI pre-coated plate	12x8 wells	FA2250	2024-06-02
TAFI standard	3 vials	FA2309	2024-05-31
Control I	1 vial	FA2306	2024-05-31
Control II	1 vial	FA2307	2024-05-31
Anti-(h)- Total TAFI HRP immunoconjugate	3 vials	FB0031	2024-06-21
Fibrinolysis Sample diluent	2x50 ml	FA1742	2025-04-21
Wash solution	1x50 ml	F2000802	2023-11-20
Conjugate diluent	1x25 ml	FA1119	2024-11-18
TMB	1x25 ml	210302D01	2025-03-30
Sulfuric Acid 0,45M	1x6 ml	F2001723	2024-06-19

Analytical Data	Specification	Result
Reactivity	A450 for 100% \geq 1.35	1.45
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.033
Detection threshold	\leq 5 %	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-02-02

Quality Control Manager : S.LECOURT