



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST PAI-1 Activity

Product Code : RK019A

Lot No. : F2001499

Expiration date : 2023-11-20

Components	Quantity	Lot	Expiration Date
tPA pre-coated plate	12x8 wells	F2001623	2024-05-13
(h) PAI-1 standard	3 vials	F2001559	2024-05-09
(h) PAI-1 Control I	1 vial	F2001560	2024-05-09
(h) PAI-1 Control II	1 vial	F2001561	2024-05-09
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	F2001645	2024-05-20
F-Sample diluent	2x50 ml	F2001621	2024-05-13
Wash solution	1x50 ml	F2000802	2023-11-20
Conjugate diluent	1x25 ml	F2001562	2024-05-08
TMB	1x25 ml	200204D01	2024-02-28
Sulfuric Acid 0,45M	1x6 ml	F2001014	2024-03-11

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml \geq 1.00	2.13
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	Controls in acceptance ranges
Blank	< 0.100	0.004
Detection threshold	\leq 0.5 ng/ml	< 0.1

Calibration Curve for information

						Acceptance Range	MV
Conc.	C	C/2	C/4	C/10	C/20	CI [5.4 - 7.2]	5.8
OD	2.518	1.569	0.876	0.378	0.183	CII [1.4 - 2.2]	1.6

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-12-22

Quality Control Manager : S.LECOURT