



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN Fibrinogen

Item Code : 120102

Lot No. : FA11271E

Expiration date : 2023-12-08

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	FA1127JA1E	2023-12-08

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	Delta A620 (0-20µg/ml) \geq 0.40 Controls measured within acceptance ranges $R^2 \geq$ 0.98	0.606 PASS PASS
Limit of Quantification	\leq 1 µg/ml (in the test dilution (1:200))	< 1 µg/mL
Hook effect	> 90 µg/ml (in the test dilution (1:200))	> 100 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-07-23

Quality Control Manager : S.LECOURT