



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN ORGARAN CONTROL  
**Product Code :** 223501  
**Lot No. :** FA117219  
**Expiration date :** 2023-12-05

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA1172CL19	2023-12-05
C2	6 vials	FA1172CM19	2023-12-05

Analytical Data	Specification	Result
Concentration	C1 : 0.40 -0.60 U/ml C2 : 0.80 -1.20 U/ml	C1 : 0.52 U/ml C2 : 1.04 U/ml

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

Quality Control Manager :

Quality Control Approval

02 JUL. 2021

S. LECOURT