



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN APIXABAN CONTROL LOW  
**Product Code :** 225201  
**Lot No. :** FA12571B  
**Expiration date :** 2023-12-19

Components	Quantity	Lot	Expiration Date
CI	6 vials	FA1257DW1B	2023-12-19
CII	6 vials	FA1257DX1B	2023-12-19

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 45 -105 ng/ml	25 ng/mL 80 ng/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-07-19

Quality Control Manager : S.LECOURT

*Plo* 

