



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

CERTIFICATE OF ANALYSIS	
<b>Product Description :</b>	<b>Factor XII Deficient Plasma</b>
<b>Product Code :</b>	<b>DP080A/K</b>
<b>Lot No. :</b>	<b>FA12471A</b>
<b>Expiration date :</b>	<b>2023-12-19</b>

Components	Quantity	Lot	Expiration Date
Factor XII Deficient Plasma	1 vial (DP080A) 6 vials (DP080K)	FA1247JY1A	2023-12-19

Analytical Data	Specification	Result
Calibration curve	CT (0%) > 180 sec CT (100%) < 65 sec	313 sec 51.5 sec
Concentration FXII	< 1%	< 1%
Concentration FV	> 50 %	81%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-07-19

Quality Control Manager : S.LECOURT


