



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PLASMA ARGATROBAN
CALIBRATOR

Product Code : SC030K

Lot No. : FA125416

Expiration date : 2023-12-16

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA1254BZ16	2023-12-16
CAL 2	4 vials	FA1254CA16	2023-12-16
CAL 3	4 vials	FA1254CB16	2023-12-16
CAL 4	4 vials	FA1254CC16	2023-12-29
CAL 5	4 vials	FA1254CD16	2023-12-16

Analytical Data	Specification	Result
CAL 1	≤ 0.05 µg/mL	0
CAL 2	0.35 - 0.65 µg/mL	0.62
CAL 3	0.85 - 1.15 µg/mL	1.12
CAL 4	1.30 -1.65 µg/mL	1.51
CAL 5	1.80 -2.10 µg/mL	2.01

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-07-23

Quality Control Manager : S.LECOURT

