



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN HEPARIN CALIBRATOR
Product Code: 222001
Lot No.: FA14621B
Expiration date: 2024-01-30

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA1462BJ1B	2024-01-30
CAL 2	4 vials	FA1462BK1B	2024-01-30
CAL 3	4 vials	FA1462BL1B	2024-01-30
CAL 4	4 vials	FA1462BM1B	2024-01-30
CAL 5	4 vials	FA1462BN1B	2024-01-30

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 0.05 IU/mL CAL 2 : 0.30 -0.50 IU/mL CAL 3 : 0.70 -0.90 IU/mL CAL 4 : 1.10 -1.30 IU/mL CAL 5 : 1.45 -1.75 IU/mL	0 IU/mL 0.36 IU/mL 0.77 IU/mL 1.15 IU/mL 1.53 IU/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2021-09-09

Quality Control Manager: S. LECOURT