



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	BIOPHEN RIVAROXABAN CALIBRATOR
Product Code :	222701
Lot No. :	FA140518
Expiration date :	2024-01-13

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA1405BA18	2024-01-13
CAL 2	4 vials	FA1405BB18	2024-01-13
CAL 3	4 vials	FA1405BC18	2024-01-13

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/mL 251 ng/mL 513 ng/mL
Linearity and accuracy of calibration with values assigned	R ² ≥ 0.98 Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-09-01

Quality Control Manager : **S.LECOURT**

