



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN APIXABAN CALIBRATOR LOW
Product Code: 226101
Lot No.: FA143619
Expiration date: 2024-01-23

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FA1436DT19	2024-01-23
CAL II	4 vials	FA1436DU19	2024-01-23
CAL III	4 vials	FA1436DV19	2024-01-23

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/mL CAL II : 40 -60 ng/mL CAL III : 80 -120 ng/mL	0 ng/mL 54 ng/mL 108 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2021-09-09

Quality Control Manager: S. LECOURT

