



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description:	BIOPHEN APIXABAN CALIBRATOR
Product Code:	226201
Lot No.:	FA138119
Expiration date:	2024-01-09

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA1381DO19	2024-01-09
CAL 2	4 vials	FA1381DP19	2024-01-09
CAL 3	4 vials	FA1381DQ19	2024-01-09

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -400 ng/ml CAL 3 : 500 -700 ng/ml	0 313 663
Linearity and accuracy of calibration with values assigned	R ² ≥ 0.98 Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2021-09-02

Quality Control Manager: S. LECOURT

