



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : RISTOCETIN
Product Code : AG004K
Lot No. : FA1441
Expiration date : 2024-01-23

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	FA1441	2024-01-23

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.5 mg/mL	≥ 80%	92%
Reactivity (% AgregMax) at 1 mg/mL	≥ 70%	84%
Reactivity (% AgregMax) at 0.5 mg/mL	< 10%	3%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-08-20

Quality Control Manager : C.DUNOIS

plp

Sophie LECOURT
Resp. CO.