



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor VII Deficient Plasma
Product Code : DP030A/K
Lot No. : FA13791C
Expiration date : 2024-01-08

Components	Quantity	Lot	Expiration Date
Factor VII Deficient Plasma	1 vial (DP030A) 6 vials (DP030K)-	FA1379JR1C	2024-01-08

Analytical Data	Specification	Result
Calibration curve	CT (0%) > 100 sec CT (100%) < 40 sec Thromboplastin : STAGO Néoptimal	117 sec 27.5 sec
Concentration FVII	< 1%	< 1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-08-12

Quality Control Manager : M.PEYRAFITTE