



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN APIXABAN CONTROL  
**Product Code :** 225301  
**Lot No. :** FA17401D  
**Expiration date :** 2024-02-13

| Components | Quantity | Lot      | Expiration Date |
|------------|----------|----------|-----------------|
| C1         | 6 vials  | FA17401D | 2024-02-13      |
| C2         | 6 vials  | FA17401D | 2024-02-13      |

| Analytical Data | Specification                              | Result                 |
|-----------------|--|------------------------|
| Concentration   | C1 : 140 -260 ng/ml<br>C2 : 340 -460 ng/ml | 193 ng/mL<br>398 ng/mL |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-10-11

Quality Control Manager : S.LECOURT

