



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description:	ZYMUTEST TOTAL TISSUE FACTOR (FULL LENGTH)
Product Code :	RK035A
Lot No. :	FA1533
Expiration date :	2024-02-16

Components	Quantity	Lot	Expiration Date
Anti-(h)-TF pre-coated plate	12x8 wells	FA1730	2024-03-02
TF Standard	3 vials	FA1694	2024-02-26
Biotinylated Antibody	1 vial	FA1724	2024-03-15
Control I	1 vial	FA1695	2024-02-26
Control II	1 vial	FA1696	2024-02-26
Assay Enhancer - TF	1x15ml	FA1611	2024-02-16
HRP streptavidine immunoconjugate	3 vials	FA1725	2024-03-18
TF-Sample diluent	2x40 ml	FA1613	2024-02-16
Wash solution	1x50 ml	FA0257	2024-07-03
TF - Conjugate diluent	1x25 ml	FA1612	2024-02-16
TMB HS	1x25 ml	200517D04	2024-05-31
Sulfuric Acid 0,45M	1x6 ml	F2001723	2024-06-19

Analytical Data	Specification	Result
Reactivity	A450 for 500 pg/ml \geq 1.60	2.34
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.026
Detection threshold	\leq 25 pg/ml	<1

Calibration Curve for information

Conc.	0	32.25	64.5	161.85	322.5	645
OD	0.026	0.208	0.358	0.829	1.622	2.725

	Acceptance Range	MV
CI	[275 - 413]	334
CII	[65 - 109]	99

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-10-11

Quality Control Manager : S.LECOURT

