



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : CEPHEN 2.5 LS
Product Code : CK522K
Lot No. : FB0578A2
Expiration date : 2024-03-31

Components	Quantity	Lot	Expiration Date
CEPHEN 2.5 LS	6 vials	FB0578KDA2	2024-03-31

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	STAR 1.05 CS 1.06

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-05-11

Quality Control Manager :

S.LECOURT