



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: ZYMUTEST Protein C

Product Code: RK027A

Lot No.: FA2089

Expiration date: 2024-03-29

Components	Quantity	Lot	Expiration Date
Anti (h) Protein C pre-coated plate	12x8 wells	FA1913	2024-03-29
Protein C Standard	3 vials	FA1975	2024-04-08
Protein C Control I	1 vial	FA1977	2024-04-08
Protein C Control II	1 vial	FA1976	2024-04-08
Anti-(h)-Protein C HRP immunoconjugate	3 vials	FA2015	2024-04-15
Sample diluent	2x50 ml	F2001717	2024-06-17
Wash solution	1x50 ml	FA2213	2025-05-12
Conjugate diluent	1x25 ml	FA1121	2024-11-18
TMB	1x25 ml	210302D01	2025-03-30
Sulfuric Acid 0,45M	1x6 ml	FA0938	2024-10-14

Analytical Data	Specification	Result
Reactivity	A450 for 100% \geq 1.40	1.59
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.010
Detection threshold	\leq 5 %	< 1 %

Calibration Curve for information

						Acceptance Range	MV	
Conc.	C	C/2	C/4	C/10	C/20	CI	[69 - 93]	80
OD	1.799	1.303	0.869	0.437	0.245	CII	[39 - 59]	47

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-02-15

Quality Control Manager: S.LECOURT

