



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description:	Factor VIII:C Deficient Plasma
Product Code:	DP040A/K
Lot No.:	FA22251F
Expiration date:	2024-05-19

Components	Quantity	Lot	Expiration Date
Factor VIII:C Deficient Plasma	1 vial (DP040A) 6 vials (DP040K)	FA2225JV1F	2024-05-19

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) \geq 40 sec CT (100%) < 75 sec	50 sec 52.3 sec
Concentration FVIII	< 1%	< 1 %
Concentration FV	> 50 %	78%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2021-12-13

Quality Control Manager: S. LECOURT

